Saelig Company, Inc.

71

info@saelig.com

Perinton Parkway, Fairport, NY 14450 USA 1-585-385-1750


www.saelig.com



NFP-3 Near Field Probe


Product Overview

NFP-3 is used with **RIGOL** DSA series spectrum analyzer for the EMI tests of electronic products. It can be used to test the magnetic field strength and magnetic field coupling channels on the surface of the electronic components as well as the magnetic field environment near the electronic module so as to quickly locate the interference source. NFP-3 includes four models (NFP-3-P1, NFP-3-P2, NFP-3-P3 and NFP-3-P4).



Measurement Connections

The connection mode of NFP-3 and spectrum analyzer is as shown in the figure below.



Connect the spectrum analyzer Connect the SMB (M) terminal of NF

Connect the SMB (M) terminal of NFP-3 and the BNC (F) terminal of the N-BNC adaptor respectively via the BNC-SMB RF cable; connect the N (M) terminal of the N-BNC adaptor to the RF input terminal of the spectrum analyzer.

RIGOL

Connect the device under test

NFP-3 is used to perform short-distance noncontact measurement on the device under test. Pay attention to the direction of the probe during measuring.

Typical Applications

- Locate the EMI radiation interference source.
- Determine the frequency and relative strength of the spectral component of the interference source.

Performance Specifications


Frequency	
Frequency Range	30 MHz to 3 GHz

Terminal Type	
Terminal Type	SMB (M)
Adaptor	N (M)-BNC (F)
RF Cable	BNC (M)-SMB (F), 1000 mm
Terminal and Adaptor Impedance	50 Ω

General Specifications	
Dimensions	260 mm × 190 mm × 30 mm (with package)
Weight	0.425 kg (with package)
Operating Temperature	0°C to 50°C
Storage Temperature	-20℃ to 70℃

Apr. 2016

Model	Description
NFP-3-P1	Near field probe for magnetic field measurements. The test range
TH TH	is within 10 cm. It is used to locate the leakage field. P1 -20 -30 -40 -50 -50 -50 -50 Frequency (MHz)
NFP-3-P2	Near field probe for magnetic field measurements. The test range
	is within 3 cm. It is used to accurately test the leakage field.
	-10 P2
	(a) -30 −
H	(gp) uigs
	-60 0 500 1000 1500 2000 2500 3000 Frequency (MHz)



Saelig Company, Inc.

71 Perinton Parkway, Fairport, NY 14450 USA 1-585-385-1750 info@saelig.com www.saelig.com